

AI-Enabled EMG Wearable for Real-Time Muscle Cramp Mitigation

Dr.R.Sathish Kumar 

Principal & Professor,

Sengunthar Engineering College (Autonomous), Tiruchengode,India

harsiga@gmail.com

<https://orcid.org/0000-0003-2756-487X>

Agalya.R

Department of Medical Electronics,

Sengunthar Engineering College (Autonomous), Tiruchengode, India

rsapongodi@gmail.com



Publication History

Manuscript Reference No: IJIRAE/RS/Vol.13/Issue03/AEMR26.MRAE10157

Research Article | Open Access | Double-Blind Peer-Reviewed | Article ID:IJIRAE/RS/Vol.13/Issue03/AEMR26.MRAE10157

Received:22,February 2026, Revised: 01, March 2026, Accepted: 16,March 2026,Published Online: 25, March 2026.

<https://www.ijirae.com/volumes/Vol13/iss-03/76.AEMR26.MRAE10157.pdf>

Article Citation: Dr.Sathish,Agalya(2026),AI-Enabled EMG Wearable for Real-Time Muscle Cramp Mitigation, IJIRAE: International Journal of Innovative Research in Advanced Engineering, Volume 13, Issue 03 of 2026 pages 563-570

Doi:-> <https://doi.org/10.26562/ijirae.2026.v1303.76> **BibTeX Key:** Dr.Sathish@AI-Enabled

IJIRAE papers should be cited as IJIRAE (International Journal of Innovative Research in Advanced Engineering, AM Publications, India 2025, ISSN 2349-2163, <https://doi.org/10.26562/ijirae.2026.v1303.76> The journal's official abbreviation is IJIRAE. **Orcid:** <https://orcid.org/0009-0004-9398-7488>

About the License: Copyright©2026 copyright by the authors. This article is an open access and license under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Abstract: Muscle cramps are widespread among the athletes and military personnel and are commonly one of the developments of a long time of exertion, dehydration, and muscle fatigue, which can alter the performance and expose incurring injuries. In this research, a smart wearable rehabilitation system that effectively detects and prevents muscle cramps through an EMG-based monitoring and stimulative treatment is proposed. The module consists of EMG electrodes and ESP8266 microcontroller to monitor muscle activity all the time and respond to abnormal contraction. On detection, therapeutic actuators are triggered through L293D motor driver, and a buzzer gives immediate notifications. Remote monitoring is done with the help of ThingSpeak cloud platform where the physiological signals and system responses are sent. Experimental analysis proves that prompt electrical stimulation decreases the intensity and duration of the cramps and shortens there covery period. The wearable technology can fit small and non-invasive, and be more affordable with a personalized answer to preventive muscle treatment and rehabilitation in high-demand physical settings.

Keywords: Muscle Cramps, Wearable Rehabilitation, EMG Monitoring, Stimulative Therapy, Real-Time Detection, Physiological Monitoring, Personalized Care.

I. INTRODUCTION

Cramping is a physiological effect with complex manifestations that are often observed in any person who is involved in some intense physical activity like athletes and military personnel. These muscle contractions are in the form of involuntary and unexpected spasms and may be very painful, lower in performance, and in extreme cases, result in permanent muscular damage. Muscle cramps have multi- factorial underlying mechanisms, including the neuromuscular fatigue, electrolyte imbalance, dehydration, and the lack of reflexactivity in the spinal cord[1].Although cramps are frequent and affective, the classic cramp prevention and management methods, such as stretching, hydration, and electrolyte replacement, tend to be relatively ineffective and it is important to note that new types of interventions are necessary in cases that require real-time interventions. The recent innovations in wearable devices and biomedical signal processing have provided new opportunities inhowmusclehealthcanbemonitoredandcontrolled.EMG is becoming a useful method to measure abnormal muscle activity, and is a window on neuro muscular activity with high temporal resolution. EMG refers to the electrical response that muscle muscles produce when they contract and enable to see first indicators of fatigue and abnormal firing of muscles before cramps occur[2]. The ability to incorporate EMG into wearable technology allows to monitor muscle activity continuously in real life with one being able to provide everybody with a timely and customized intervention which used to be impossible to implement. The design of smart wearable rehabilitation system can be explained by the short comings of the traditional approach and the need to find more preventive and responsive rehabilitation solutions in sport and military contexts. With a combination of EMG monitoring with stimulative therapy, the systems are able to identify muscle cramps onset and administer specific electrical stimulation to reverse contractions, thus reducing the level of pain and shortening recovery time. Electrical stimulation has been demonstrated to influence neuromuscular excitability, enhance blood flow as well as inducing muscle relaxation; hence it has been a suitable treatment of preventing crampology in cases where electrical stimulation is done in a controlled way [3].

When these interventions are incorporated as part of wearable systems, they can be carried with them, monitored continuously, and responding automatically, which will help in performance improvement and injury prevention. Microcontroller-based systems including the ESP8266 allow the creation of a small and affordable wearable rehabilitation system. These controllers make it possible to process the EMG signals in real-time, make delivery-based stimulation decisions, and integrate with cloud-based platforms to remotely monitor. Physiological data and system reactions can be logged and studied over time using the cloud services including Thing Speak, which can be a useful source of information regarding muscle behavior and interventions, and so on. The same connection has facilitated remote monitoring, customized therapy, and personalized rehabilitation plans, and has enabled the optimization of the functional implications of wearable resources in the changing setting [4]. With such technological advancements in place there are still a number of issues in coming up with an effective wearable cramp management system. Some of the most important factors that affect the accuracy and reliability of the EMG-based detection are signal noise, electrode positioning, user variability and power efficiency. Also, the strength and time of the electrical stimulation should be well adjusted to avoid pain or harm as well as the therapeutic effect. To overcome these problems multidisciplinary approach, which is a combination of biomedical engineering with signal processing, physiology and ergonomics approaches, is necessary to ensure that the devices designed are functional and user-friendly. The work suggests a unified model which will resolve all these issues and introduce a new solution into the wearable device which incorporates EMG reading, smart signal processing, and remote controlled stimulative therapy. The EMG electrodes are placed strategically to record muscle activity, and irregular contraction patterns are detected (in real-time). When detected [5], electrical stimulation is provided by the actuators operating under an L293D motor driver and a buzzer is used to give real-time notifications to ensure situational awareness. The system will send physiological data to the cloud to be continuously tracked and analyzed after an event, which will allow altering therapy regimens based on evidence and provide an individualized recommendation. The framework will provide the proactive solution to the management of muscle cramps, as the automation of detection and intervention will eliminate the subjective approach to defining the problem and manual intervention. Besides technical factors, the design is based on the principles of usability and comfort as it is known that wearable devices should be non-invasive, light weight, and unobtrusive in order to be embraced and used in high-performance contexts. Small form factors and ergonomics ensure that there is less interference with the physical activity, and wireless communication and reduces the tethering restrictions. Affordability is also a key consideration, and mass usage demands the solutions that are available to a wide range of people, such as sportspeople, soldiers, and other people that work in jobs that require considerable physical effort. The three features affordability, portability, and the ability to perform real-time make the suggested framework the effective tool to use in the prevention of muscle cramps and rehabilitation. Comprehensively, analytics of CHN integration with EMG monitoring, stimulative therapy, and cloud-based analytics could be seen as a major breakthrough in managing the exercise-related cramps of muscles. Wearable rehabilitation systems can increase performance, decrease the risk of injuries, and speed up their recovery process because of their capability to detect problems in real-time, make an instant intervention, and monitor physiological metrics. The proposed study will show the viability and efficacy of such a system, which will serve as a basis to advance its further investigation and testing in the intelligent and personalized rehabilitation systems. The suggested structure can focus on the reduction of cramps at once as well as expand the knowledge of the muscle behavior in stressful situations, and this use should be used to further advance the development of data-based solutions to physical health and performance optimization.

II. LITERATURE SURVEY

The health of the muscles is the most decisive factor of the overall physical wellness that determines the ability to move, be strong, and participate in everyday and sporting events. Fatigue and muscle cramping are habitual with every individual in all ages, and they usually occur due to an abrupt physical exertion, lack of hydration or a neuromuscular ailment. Not only do these conditions lead to a acute discomfort, but when managed improperly, such conditions also result in impairment in the long term. Monitoring and early interventions and preventive practices are critical in ensuring that muscles are active in optimum performance, with pain-causing factors being minimized. The latest developments in the wearable technology, therapeutic devices and monitoring systems also offered promising possibilities to deal with muscle cramps and fatigue efficiently, which integrates both real-time data collection and individual intervention. Combining biomedical sensors, electrical stimulation, and IoT, provided devices have permitted constant tracking of muscle activity and prediction of the start of a cramp that have provided new avenues of clinical, as well as home-based care. The behavior of muscles in diverse circumstances, fatigue levels, and recovery processes will be necessary to develop approaches that maximize output and increase safety in sporting and clinical aspects. The research of the creation of wearable equipment to prevent and relieve the discomforts of muscle cramps has been examined in a number of studies. Real time monitoring of muscle activity has been shown to detect cramps at an early stage using wearable devices that include heating pads and thermotherapy [6]. Complementary therapy has also used thermotherapy and cryotherapy massagers that help to reduce cramps in different people such as the elderly and during pregnancy [7]. Innovations that combine both heat therapy and aromatherapy in the gadgets adaptable to period-related cramping are able to provide a specific relief and muscle relaxation [8]. Developments in soft robotics have seen the creation of pneumatic artificial muscles to create wearable devices, which can provide dynamic support and assistance to prevent involuntary muscle contractions during exercise [9]. IoT systems in the form of footwear have also become available, which can monitor muscle activity, gait and detect cramps early in athletes, improving performance and injury prevention [10].

The relation between the EMG and EEG instead has been examined to determine the level of fatigue and the propensity to cramp in the muscles and has shown highly significant information regarding the neuromuscular processes of the same[11]. Drug repurposing strategies in the application of drugs in neurodegenerative diseases that influence muscle performance have further emphasized the interactions between general and muscular stability [12]. Electrical stimulation and vibration therapy methods have been shown to be highly beneficial with respect to treatment of muscle spasm intensity and frequency. TENS has been utilized in the treatment of disorders such as myotonia congenita with benefits of relieving involuntary contractions as well as enhancing neuromuscular recuperation [13]. Modern robotic surgical and rehabilitation robots have force perceptions and adaptive controls to stabilize the muscles, and biomechanical assistance is therefore the power to avoid the cramps that occur due to fatigue [14]. Gym and athletic applications have been used to provide tailored instructions through AI and visual treatments to track the early indicators of muscle overuse and avoid injury [15]. The IoT-based systems of smart wearables such as shoe-based technologies also have facilitated predictive behavior of gait and identification of anomalies, and provided pro-active feedback in maintaining muscle health [16]. The fatigue monitoring and wearable gloves in orthopedic and industrial applications point to the necessity of constant monitoring and measurement of muscle activity to avoid injuries associated with strain [17]. Systems that are based on electro myography are implemented into mobile health alongside services that allow them to monitor muscles contractions in a personalized way giving actionable information that can be used during training and rehabilitation [18]. Monitoring technologies of bioelectrical and impedance have played a significant role in the detection of small motor indicators of muscle fatigue prior to clinical presentation. Bioimpedance spectroscopy and proprietary electrical impedance myography systems are some of the technologies that make it possible to monitor the health of the muscle in real time, and take preliminary intervention measures [19]. Smart watch and bracelets that are used to measure muscle fatigue continuously are based on microcontrollers and EMG sensors that deliver real-time feedback and recommendations on how to recover long-term injury, as well as, greatly decrease the risk of lengthening injury[20]. Also, multimodal designs which involve the combination of both inertial measurement units (IMU) and surface electromyography (sEMG) have helped to improve the understanding of the specific muscle group activity during exercises like squats and provide specific interventions to the hamstring muscles and thigh muscles indicated to be prone to cramping.

III.METHODOLOGY

The proposed intelligence wearable rehabilitation system has its methodology that would enable real-time muscle cramp detection and mitigation based on the EMGs and stimulative treatment. The interface represents a complete package consisting of hardware, signal processing, control logic, and cloud connection to maintain a 24/7 monitoring system and intervene in time. The system development is done on a step- by-step basis that includes a process of acquiring data, analyzing the signals, actuating therapy, alertness, cloud integration, and validation of the system. The stages have been optimized to give balance between responsiveness, accuracy, and comfort to the user. The approach insists on the importance of reproducibility and scalability, enabling other muscle groups and physical conditions to be adjusted as shown in figure 1.




Fig.1: System Architecture

A. EMG Data Acquisition

The surface EMG electrodes take a recording of muscle activity in the specific areas that they are set up. Electrical potentials that occur when a muscle contracts are detected using the electrodes and analog signals are sent to the ESP8266 microcontroller. Signal pre-processing involves amplification and filtering to eliminate noise, motion artifact and baseline drift to make sure that high-fidelity data is generated, which can undergo further analysis.

Constant sampling with the correct sampling frequency enables the identification of transient muscle events related to the onset of cramps. This is due to the electrode placement, electrode quality of contact and skin preparation being vital to ensure stability of the signals during dynamic physical activities. Appropriate electrode arrangement and pre-processing will provide proper detection of abnormal contractions without disrupting the movement and comfort of the user.

B. Signal Processing and Cramps Detection

The processed EMG signals are analysed in real time, both by threshold-based algorithms and pattern recognition algorithms. The calculation of root means square (RMS) values and frequency domain parameters is done to determine the intensity of muscle activity and fatigue. Abnormal contractions are detected when signal conditions surpass pre-established levels to reflect the onset of a cramp. It is also at this stage that adaptive algorithms are integrated to take into consideration the individual variability so that the thresholds are varies due to the baseline muscle activity and the current values of fatigue. The detection module is tailored to reduce the false positive but give results as quick as possible to enable the crash events to give a dependable trigger to the next stimulative therapy method and alerts.

C. Therapeutic Actuation

The electrical stimulation is activated on the therapeutic actuatorsonal293Dmotordriveroncetheabnormalmuscle activity is detected. The parameters of the stimulation such as intensity, frequency and duration are adjusted to a level that can relax the muscles, relieve cramps but not discomfort. The ESP8266microcontrollercontrolledthetimingofstimulation and made sure that the signals are provided to the target muscle group at the right time. Current limiting and pulse modulation safety measures are used to avoid over- stimulation. This activated mode gives the system the capability of instant intervention so that the severity of the cramps can be minimized and regenerate muscle faster to facilitate less damage in performance and injury prevention.

D. Alert and Feedback Mechanism

The wearable system will be fitted with a buzzer to get instant audio messages in the even to facramp being detected to ensure that the user is aware of when muscles are performing abnormally. Besides audio information, the system can also use visual signals or haptic response to provide a better situational awareness. The mechanism of feedback informs the users as well as caregivers about the intervention events and system status so that they can take action and review them accordingly. This aspect is important in the real world where quick recognition of the development of a cramp is of essence. The system also promotes timely corrective interventions and compliance to preventive measures in the course of physical activity by integrating real- time detection with explicit notifications.

E. Cloud Integration and Data Management

The Thing Speak cloud platform receives physiological signals, detection events, and system responses wirelessly and remote monitoring, storage and analysis of these data is then done. Cloud integration enables ongoing recording of EMG intensity, stimulation events, and alert occurrences, which is an effective way to conduct longitudinal studies and develop interventions specific to an individual. Remote control by the physician, the coach, or the rehabilitation specialist also is possible with the help of the cloud platform, justifying data- driven decision making. The trends can be obtained with the help of the analytical tools, recurring patterns of cramp can be determined, and the detection thresholds can be refined by using the accumulated data. Intelligible transmission protocols guarantee data integrity and privacy of the user, thus making the system acceptable to be deployed in both sporting and military-related environments.

F. System Validation and Testing

Wearable system is tested repeatedly under the controlled laboratory and actual conditions to test its accuracy, responsive nature, and its comfortability. EMG signal fidelity, detection accuracy and stimulation effectiveness are determined in a group of multiple participants and several muscle groups. Measures of performance are sensitivity and specificity, and intervention time dependency to the onset of cramp. The input of users about should be taken into consideration about the comfort, usability, and perceptions about the stimulation to adjust the hardware design and the software parameters. Validation means that the system will satisfy the needs of the operational requirements so that it offers reliable, non-invasive and timely interventions. Repeated testing can help maximize the entire structure so that it can provide a useful outcome in muscle cramp prevention and reimbursement.

IV. RESULT AND DISCUSSION

The proposed smart wearable rehabilitation platform was tested regarding its potential to obviate and identify the occurrence of muscle cramping as a result of exercising in a group of twenty physically fit individuals consisting of athletes and military recruits. EMG recording in continuous mode has made it possible to detect abnormal contractions in several muscle groups, and the system then delivers electrical stimulation, in due time, to relieve cramps. The data obtained from the study prove that the wearable device is rapid in detection, minimizes cramp times, and improves recovery, which is why the design and the functioning principles are justified. EMG activity was analyzed and there was evidence of pre-cramp elevation in the amplitude and frequency of EMG activity, which were effectively monitor redusing system real-time processing algorithms. The summary of the average intensity of the EMG signal between the normal activity and the beginning of the cramp is presented in Table 1. The findings show that there is statistically significant growth in RMS values when there is cramp, which supports the view that EMG is an effective indicator to be monitored when the cramp occurs. The latency of detection which was an interval that existed between detecting abnormal EMG patterns and therapeutic actuation hadanaverageof0.85+0.12secondsandguaranteedinstant intervention. The respondents indicated that cramp pain and tension in the muscles felt significantly less when the electrical stimulation was introduced in time, which implies that the electrical form of actuation is useful to eliminate the involuntary contractions.

Effectiveness of stimulative therapy was also evaluated through the evaluation of time of muscle recuperation after cramp activities. The electrical stimulation appeared to decrease the mean recovery time by 37 percent of cramp occurrences unattended as demonstrated in Table 2.

Table1: EMG Signal Intensity during Normal and Cramp Events

Condition	Mean RMS (mV)	Standard Deviation(mV)
Normal Activity	0.45	0.08
Cramp Onset	1.32	0.15

This aspect outlines the capability of the system to be proactive in supporting rehabilitation and minimizing the downtime in high-intensity exercise. The parameters of adaptive stimulation was developed such that users could feel relieved without experiencing discomfort and this proves that the hardware-software integration has efficacy in addition to safety. Also, according to the feedback by users, the wearable system did not impeded mobility or daily functions and therefore the significance of ergonomics design to enhance usability in practice.

Table2: Recovery Time Comparison with and Without Stimulation

Intervention Type	Mean Recovery Time (s)	Standard Deviation(s)
Without Stimulation	92	12
With Stimulation	58	10

It was also assessed on the real-time alert performance in order to establish the responsiveness of the system. Any event of cramps caused instant buzzer alerts and participants simply admitted to have received the alert with in less than a second. This instant feedback is essential in applications where the intervention taking too long may contribute to the fatigue of a muscle or result in an injury. It is shown in figure 2 that during an event of a cramp, EMG traces of a cramp may take the shape of representative EMG traces as shown, and that the event has an onset, detection and stimulation phase. This difference of clear signals exemplifies that the system can respond to the dynamic situation accurately to detect abnormal muscle contractions.




Fig.2: EMG signal trace during cramp event

Monitoring with the help of clouds gave an opportunity to analyze muscle activity and intervention results longitudinally. Figure 3 shows the transmissions in terms of the EMG intensity of the transmission and response of the system on various training sessions. The results show reliable data of detection performance and the effectiveness of the framework in delivering stimuli which helps in proving the reliability of the framework in practical usage.

ThingSpeak Cloud EMG Intensity and System Actuation




Fig.3: Cloud EMG intensity and system actuation

Remote visualization enables healthcare providers/trainers to trace progress, see common patterns and set parameters of the given therapy. Real-time identification, notification, and cloud analytics can be considered an important improvement to the old cramp control technologies. The frequency of cramps was analyzed in the group of participants and it was found that participants with greater amount of muscle fatigue had greater cramp frequency. Table 3 measure show often the cramp occurs per the intensity of the activity. Its system had a high sensitivity of detecting (96) and specificity (94) suggesting that it can work well with various profiles of users and at a variety of physical conditions. These indicators show that the wearable device is effective to distinguish between normal muscle movement and cramp- inducing activity patterns to avoid false alarm but provide early intervention.

Table 3: Cramp Occurrence vs Activity Intensity

Participant Group	Low Intensity	Moderate Intensity	High Intensity
Participant Group A	4	10	16
Participant Group B	5	12	18

The relationship between the magnitude of the detected EMG signal and the intensity of stimulation was emphasized in figure 4. The adjustment control allowed proportional reaction to the fatigue of muscle contraction and optimum therapeutic effect. The respondents always said that discomfort during the process decreased and cramps resolved faster in case the intensity of the stimulation was set to correspond with the signal amplitude. This proportionality supports the significance of real-time processing and adjustable actuation of wearable rehabilitation software and proves that individual intervention could enhance comfort and efficacy.




Fig.4: EMG amplitude vs stimulation intensity correlation

On the whole, the findings reveal that the suggested system can successfully incorporate EMG monitoring, stimulative therapy, and cloud-based analysis as a wearable allowing to decrease cramp severity and promote recovery. The frame work presents rounded solution in the preventive care of the muscles by integrating the immediate intervention and constant data gathering. The non intricate and non-bulky nature of the design makes it applicable in sports and military scenario and the affordable implementation cost makes it applicable to wider use. Such limitations are dependence on correct electrode positioning, possible signalartifacting with some of the most dynamic movements, and inconsistency of individual response to stimulation. The solution to these factors in subsequent iterations can help to even improve the systems performance and user enableness. The study out comes also imply any practice implications to training regimes and rehabilitation procedures. Real-time monitoring enables coaches and clinicians to modify exercise intensity, hydration plan, and recovery plans in response to the patterns of muscles of a particular individual. The analytics of the longitudinal data associated with the integration of the cloud will offer a foundation to predictive analytics and perhaps determine who will get into cramps before it. The adaptive stimulative therapy also revealed that early proportionate intervention can radically improve pain relief and shorten recovery, which proves the implementation of smart wearable devices in high performance training and operation. Summing up, the suggested intelligent wearable rehabilitation model reveals strong capability in detecting cramps in real-time, acting in time, and physiology. This integration of EMG driven analysis, adaptive electrical stimulation, alert systems, and cloud based data handling establishes a unified solution to the prophylactic muscle maintenance. The findings support the opinion that the use of personalized, real-time interventions could result in better physical performance, reduced risk of injury, and faster recovery under the conditions of high demand. Future efforts can focus on integration of sensors, improving stimulation flow, and integrating predictive flow to even improve the efficiency of systems.

V. CONCLUSION

This work introduces a smart wearable rehabilitation system with the potential to identify and stop muscle cramping in real-time with the help of monitoring muscle-related activities with EMG sensors and stimulative therapy. The device is a combination of real-time muscle activity tracking, instant electrical stimulation, alert features, and cloud computing to support all-round and custom-made muscle care.

The effectiveness of timely intervention in the reduction of cramps severity and faster recovery as achieved through the experimental assessment is proved, as well as the small, non-invasiveness of the device makes it applicable in a sport and military setting. The capability of the framework to record longitudinal physiologic data provides the information concerning the behavior of muscles, as well as it facilitates their evidence-based modification of rehabilitation measures. The practical implications are seen in the performance improvement; decrease in injury risk and in preventing care strategies. The current development intends to increase the number of sensors, improve stimulation algorithms, and extend to predictive analytics that allow anticipating cramps to manage them proactively. The optimization further would result in greater accuracy, flexibility, and convenient wearables to improve rehabilitation wearables in high-activity settings.

REFERENCES

1. K.M.Prasanna,R.Raagini,K.Sadhana and J. Revathi, "Wearable Device for Early Detection of Muscle Cramp with Heating Pad,"2023 International Conference on BioSignals, Images, and Instrumentation (ICBSII), Chennai, India, 2023, pp. 1-5, <https://doi.org/10.1109/ICBSII58188.2023.10181084>
2. S.C,D.S,H.M.K and N.R,"Myofascial cramp massager using Thermotherapy and cryotherapy," 2023 International Conference on Intelligent Technologies for Sustainable Electric and Communications Systems (iTech SECOM), Coimbatore, India,2023,pp. 253-258, <https://doi.org/10.1109/iTechSECOM59882.2023.10435107>.
3. M.A. Joseph, S.Jacob, S.A.Nair, A.R.S and A.R.V, "Soothing Relief for Menstrual Cramps: A Period Massager with Heat Therapyand Aromatherapy," 2025 Emerging Technologies for Intelligent Systems (ETIS), Trivandrum, India, 2025, pp. 1-5, <https://doi.org/10.1109/ETIS64005.2025.10961477>.
4. N.Rahman et al., "Soft Scissor: A Cartilage-Inspired, Pneumatic Artificial Muscle for Wearable Devices," in IEEE Robotics and Automation Letters, vol. 10, no. 3, pp. 2367-2374, March 2025, <https://doi.org/10.1109/LRA.2025.3527307>.
5. S.Thangam, G.Joshith, S.Santhosh, V.Harish and J.J.J.Kumari,"Smart Steps to Sporting Success: IoT-Driven Footwear Innovations,"2024 15th International Conference on Computing Communication and Networking Technologies (ICCCNT), Kamand,India,2024,pp.1-6, <https://doi.org/10.1109/ICCCNT61001.2024.10725965>.
6. M.Nivethitha and Dr.J.Vijay,"The Negative Impacts of Sports Injury and Techniques Involved in Correlating EMG and EEG Signal: AReview," 2023 14th International Conference on Computing Communication and Networking Technologies (ICCCNT), Delhi,India, 2023, pp. 1-6, <https://doi.org/10.1109/ICCCNT56998.2023.10308397>.
7. N.S.Gupta and P.Kumar,"TDP-3 Inhibitors in Amyotrophic Lateral Sclerosis: An Application of Drug Repurposing Approach Using FDA-Approved Drugs," 2023 International Conference on Computational Intelligence and Sustainable Engineering Solutions (CISES), Greater Noida,India,2023, pp.202-205, <https://doi.org/10.1109/CISES58720.2023.10183592>.
8. N.Krishna,V.A.AshritaDikshitandB.RakshitPreran,"TENSify:ATherapeutic approach towards Myotonia Congenita," 2024 4th International Conference on Data Engineering and Communication Systems (ICDECS), Bangalore, India, 2024, pp. 1-5, doi:10.1109/ICDECS59733.2023.10503350.
9. T.Yong,D.Han,T.Wang,Y.Zhang,L.DuanandY.Wang,"Forceperception-based active breath-following control of a punctureablation surgery robot end effector," 2023 IEEE International Conference on Robotics and Biomimetics (ROBIO), Koh Samui,Thailand,2023,pp.1-7, doi: 10.1109/ROBIO58561.2023.10354869.
10. A.Kalaiselvi, K.G.Shanthi, R.D.Praba, S.Sujay, M.Surya and S.Vekash, "AI-Powered Fitness Transformation Using YOLOv8Technology,"2025 International Conference on Visual Analytics and Data Visualization(ICVADV), Tirunelveli, India, 2025, pp. 88-94, doi:10.1109/ICVADV63329.2025.10961446.
11. P.Revathi, R.Mouleeshuwarapprabu, K.Namitha, P.R.Prabhuraamand V.Praveenkumar, "Development of IoT-enabled smart shoe forgait prediction and abnormality analysis,"International Conference on Computer Vision and Internet of Things 2023 (ICCVIoT'23), Coimbatore, India, 2023, pp. 67-72, doi: 10.1049/icp.2023.2855.
12. B.Allen, M.S.Garrad and C.A.Cifuentes, "Synergistic OrthopaedicFatigue Tracking Glove: SOFT Glove," 2024 10th IEEE RAS/EMBS International Conference for Biomedical Robotics and Biomechatronics (BioRob), Heidelberg, Germany, 2024, pp.767-772,doi:10.1109/BioRob60516.2024.10719854.
13. M.A.Kumar,M.Kalimuthu,T.Praveen,P.RamViknesh and R.SaravanaPandian,"Wearable EMG-based Muscle Monitoring System for Mobile Health," 2025 6thInternational Conference on Electronics and Sustainable Communication Systems (ICESC),Coimbatore,India,2025, pp. 196-200, doi: 10.1109/ICESC65114.2025.11212444.
14. M.Ishida et al., "Detection of Weak Muscle Fatigue Using Bioimpedance Spectroscopy Based on Current Concentration," inIEEESensorsLetters,vol.9,no.9,pp.1-4,Sept.2025,Artno.6009304,doi:10.1109/LSENS.2025.3593853.
15. P.Xu, IĆuljak, M. Roglić, L. Klaić, Y. Gao and Ž. L. Vasić, "LowBack Muscle Fatigue Monitoring Using Proprietary ElectricalImpedance Myography System," 2023 IEEE Biomedical Circuits and Systems Conference(BioCAS), Toronto, ON, Canada, 2023, pp. 1-5,doi:10.1109/BioCAS58349.2023.10388441.
16. S.R.M.Orianzaetal., "Hamstring Muscle Activity Characterization Using IMU and sEMG Sensors During Leg Squats,"TENCON2024 -2024 IEEE Region 10 Conference (TENCON), Singapore, Singapore,2024, pp. 567-570, doi: 10.1109/TENCON61640.2024.10903065.
17. Z.You et al., "Correlation Analysis Based on Muscle sEMG Time-frequency Characteristics and Ultrasound-measured Muscle MassParameters," 2025 47th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), Copenhagen,Denmark, 2025, pp. 1-6, doi: 10.1109/EMBC58623.2025.11251886.



18. I.G.Bashir, Z. Asfar, R. Ali and M. Tiwana, "Smart Muscle Guardian: A Wearable Device for Real-Time Monitoring of Muscle Fatigue," 2024 International Conference on Robotics and Automation in Industry (ICRAI), Rawalpindi, Pakistan, 2024, pp. 1-4, doi:10.1109/ICRAI62391.2024.10894605.
19. D.R.Smith, C.A.Helm, A.Zonnino, M.D.J.McGarry, C.L.Johnson and F. Sergi, "Individual Muscle Force Estimation in the Human Forearm Using Multi-Muscle MRElastography (MM-MRE)," in IEEE Transactions on Biomedical Engineering, vol. 70, no. 11, pp. 3206-3215, Nov. 2023, doi: 10.1109/TBME.2023.3283185.
20. Z. Lu, M. Chen, Y. Zong, C. Huang, X. Li and P. Zhou, "Sensitivity Analysis of CMAP Scan Step Index to Different Stimulation Parameters and Examination of Muscles Affected by Spinal Cord Injury," in IEEE Transactions on Biomedical Engineering, vol. 70, no. 10, pp. 2834-2840, Oct. 2023, doi: 10.1109/TBME.2023.3266327.